

# AmpliSens® *Mycoplasma hominis*-FRT PCR kit



For Professional Use Only

## Instruction Manual

### KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Contains sufficient for <n> tests
	<i>In vitro</i> diagnostic medical device		Use-by Date
	Version		Consult instructions for use
	Temperature limit		Keep away from sunlight
	Manufacturer	<b>NCA</b>	Negative control of amplification
	Date of manufacture	<b>C-</b>	Negative control of extraction
	Authorized representative in the European Community	<b>C+</b>	Positive control of amplification
		<b>IC</b>	Internal control

### 1. INTENDED USE

AmpliSens® *Mycoplasma hominis*-FRT PCR kit is an *in vitro* nucleic acid amplification test for qualitative detection of *Mycoplasma hominis* DNA in the clinical materials (urogenital swabs, urine, and prostate gland secretion) using real-time hybridization-fluorescence detection of amplified products.

**NOTE:** The results of PCR analysis are taken into account in complex diagnostics of disease.

### 2. PRINCIPLE OF PCR DETECTION

*Mycoplasma hominis* DNA detection by the polymerase chain reaction (PCR) is based on the amplification of the pathogen genome specific region using specific *Mycoplasma hominis* primers. In the real-time PCR, the amplified product is detected with the use of fluorescent dyes. These dyes are linked to oligonucleotide probes which bind specifically to the amplified product during thermocycling. The real-time monitoring of fluorescence intensities during the real-time PCR allows the detection of accumulating product without re-opening the reaction tubes after the PCR run.

AmpliSens® *Mycoplasma hominis*-FRT PCR kit is a qualitative test that contains the Internal Control (Internal Control-FL (IC)). It must be used in the extraction procedure in order to control the extraction process of each individual sample and to identify possible reaction inhibition.

AmpliSens® *Mycoplasma hominis*-FRT PCR kit uses "hot-start", which greatly reduces the frequency of nonspecifically primed reactions. In variant FRT, "hot-start" is guaranteed by the separation of nucleotides and Taq-polymerase using a wax layer. Wax melts and reaction components mix only at 95 °C. In variant FRT-100 F, "hot-start" is guaranteed by using chemically modified polymerase (TaqF). The chemically modified polymerase (TaqF) is activated by heating at 95 °C for 15 min.

The PCR kit contains the system for prevention of contamination by amplicons using the enzyme uracil-DNA-glycosylase (UDG) and dUTP. The enzyme UDG recognizes and catalyzes the destruction of the DNA containing deoxyuridine, but has no effect on DNA containing deoxythymidine. Deoxyuridine is absent in the authentic DNA, but is always present in amplicons, because dUTP is a part of dNTP mixture in the reagents for the amplification. Due to the deoxyuridine containing contaminating amplicons are sensitive to the destruction by UDG before the DNA-target amplification. So the amplicons cannot be amplified.

The enzyme UDG is thermolabile. It is inactivated by heating at temperature above 50 °C. Therefore, UDG does not destroy the target amplicons which are accumulated during PCR. The results of amplification are registered in the following fluorescence channels:

Table 1

Channel for fluorophore	FAM	JOE
DNA-target	<i>Mycoplasma hominis</i> DNA	Internal Control-FL (IC) DNA
Target gene	<i>gene</i> 16S rRNA	Artificially synthesized sequence

### 3. CONTENT

AmpliSens® *Mycoplasma hominis*-FRT PCR kit is produced in 2 forms:

variant FRT R-B3(RG)-CE;

variant FRT-100 F R-B3-F(RG,iQ)-CE.

Variant FRT includes:

Reagent	Description	Volume, ml	Quantity
PCR-mix-1-FL <i>Mycoplasma hominis</i> ready-to-use single-dose test tubes (under wax)	clear liquid from colorless to light lilac colour	0.01	110 tubes of 0.5 or 0.2 ml
PCR-mix-2-FL-red	red clear liquid	1.1	1 tube
Positive Control complex (C+)	colorless clear liquid	0.2	1 tube
DNA-buffer	colorless clear liquid	0.5	1 tube
Negative Control (C-)*	colorless clear liquid	1.2	1 tube
Internal Control-FL (IC)**	colorless clear liquid	1.0	1 tube

\* must be used in the extraction procedure as Negative Control of Extraction.

\*\* add 10 µl of Internal Control-FL (IC) during the DNA extraction directly to the sample/lysis mixture (see DNA-sorb-AM protocol).

AmpliSens® *Mycoplasma hominis*-FRT PCR kit is intended for 110 reactions (including controls).

Variant FRT-100 F includes:

Reagent	Description	Volume, ml	Quantity
PCR-mix-1-FL <i>Mycoplasma hominis</i>	clear liquid from colorless to light lilac colour	1.2	1 tube
PCR-mix-2-FRT	colorless clear liquid	0.3	2 tubes
Polymerase (TaqF)	colorless clear liquid	0.03	2 tubes
Positive Control complex (C+)	colorless clear liquid	0.2	1 tube
DNA-buffer	colorless clear liquid	0.5	1 tube
Negative Control (C-)*	colorless clear liquid	1.2	1 tube
Internal Control-FL (IC)**	colorless clear liquid	1.0	1 tube

\* must be used in the extraction procedure as Negative Control of Extraction.

\*\* add 10 µl of Internal Control-FL (IC) during the DNA extraction directly to the sample/lysis mixture (see the DNA-sorb-AM protocol).

Variant FRT-100 F is intended for 110 reactions (including controls).

### 4. ADDITIONAL REQUIREMENTS

- Transport medium.
- DNA extraction kit.
- Disposable powder-free gloves and laboratory coat.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol filters (up to 100 µl).
- Tube racks.
- Vortex mixer.
- Desktop centrifuge with a rotor for 2-ml reaction tubes.
- PCR box.
- Real-time instruments (for example, Rotor-Gene 3000/6000 (Corbett Research, Australia); iCycler iQ or iCycler iQ5 (Bio-Rad, USA), Mx3000P (Stratagene, USA) or equivalent).
- Disposable polypropylene tubes when working with PCR kit variant FRT-100 F:
  - a) thin-walled 0.2-ml PCR tubes with domed caps if a plate-type instrument is used;
  - b) thin-walled 0.2-ml PCR tubes with flat caps or strips of four 0.1-ml Rotor-Gene PCR tubes if a rotor-type instrument is used.
- Refrigerator for 2–8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir for used tips.

## 5. GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol filters and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distinctly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all samples or reagents spills using a disinfectant, such as 0.5 % sodium hypochlorite or another suitable disinfectant.
- Avoid inhalation of vapors, samples and reagents contact with the skin, eyes, and mucous membranes. Harmful if swallowed. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice if necessary.
- Safety Data Sheets (SDS) are available on request.
- Use of this product should be limited to personnel trained in DNA amplification techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents in the area where the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

## 6. SAMPLING AND HANDLING

Obtaining samples of biological materials for PCR-analysis, transportation and storage are described in manufacturer's handbook [1]. It is recommended that this handbook is read before starting work.

**AmpliSens® Mycoplasma hominis-FRT** PCR kit is intended for analysis of the DNA extracted with DNA extraction kits from the clinical material (urogenital swabs, urine samples (sediment of the first portion of the morning specimen), prostate gland secretion).

## 7. WORKING CONDITIONS

**AmpliSens® Mycoplasma hominis-FRT** PCR kit should be used at 18–25 °C.

## 8. PROTOCOL

### 8.1. DNA extraction

It is recommended that the following nucleic acid extraction kits are used:

- **DNA-sorb-AM.**
- For other nucleic acid extraction kits see Guidelines [2].

The DNA extraction of each test sample is carried out in the presence of **Internal Control-FL (IC)**.

In the extraction procedure it is necessary to carry out the control reactions as follows:

- C– – Add **100 µl of Negative Control (C–)** to the tube labeled C–.

**NOTE:** Extract DNA according to the manufacturer's protocol.

### 8.2. Preparing PCR

#### 8.2.1 Preparing tubes for PCR

The type of tubes depends on the PCR instrument used for analysis. Use disposable filter tips for adding reagents, DNA and control samples into tubes.

##### Variant FRT

The total reaction volume is **30 µl**, the volume of DNA sample is **10 µl**.

1. Prepare the required number of the tubes with **PCR-mix-1-FL Mycoplasma hominis** and wax for amplification of DNA from clinical and control samples.
2. Add **10 µl of PCR-mix-2-FL-red** to the surface of the wax layer into each tube ensuring that it does not fall under the wax and mix with **PCR-mix-1-FL Mycoplasma hominis**.

##### Variant FRT-100 F

The total reaction volume is **25 µl**, the volume of DNA sample is **10 µl**.

1. Thaw the tube with **PCR-mix-2-FRT**. Vortex the tubes with **PCR-mix-1-FL Mycoplasma hominis**, **PCR-mix-2-FRT**, and **polymerase (TaqF)** then centrifuge briefly. Take the required number of the tubes/stripes for amplification of the DNA obtained from clinical and control samples.
2. For N reactions (including 2 controls of amplification), add to a new tube:
  - 10\*(N+1) µl of PCR-mix-1-FL Mycoplasma hominis;**
  - 5.0\*(N+1) µl of PCR-mix-2-FRT;**
  - 0.5\*(N+1) µl of polymerase (TaqF).**
 Vortex the tube, then centrifuge briefly. Transfer **15 µl** of the prepared mixture to each tube.

Steps 3 and 4 are required in both variants.

- 3 Add **10 µl** of **DNA** obtained at the DNA extraction stage.

- 4 Carry out the control reactions:

- NCA** – Add **10 µl** of **DNA-buffer** to the tube labeled **NCA** (Negative Control of Amplification).
- C+** – Add **10 µl** of **Positive Control complex (C+)** to the tube labeled **C+** (Positive Control of Amplification).
- C–** – Add **10 µl** of the **sample extracted from the Negative Control (C–) reagent** to the tube labeled **C–** (Negative control of Extraction).

## 8.2.2. Amplification

1. Create a temperature profile on your instrument as follows:

Table 2

Step	Rotor-type Instruments <sup>1</sup>			Plate-type Instruments <sup>2</sup>		
	Temperature, °C	Time	Cycles	Temperature, °C	Time	Cycles
1	95	15 min	1	95	15 min	1
2	95	5 s	5	95	5 s	5
	60	20 s		60	20 s	
	72	15 s		72	15 s	
3	95	5 s	40	95	5 s	40
	60	20 s		60	30 s	
		Fluorescence acquiring			Fluorescence acquiring	
72	15 s	72	15 s			

Fluorescent signal is detected in the channels for the FAM and JOE fluorophores (other channels are enabled if several tests are simultaneously carried out in a single run).

2. Adjust the fluorescence channel sensitivity according to *Important Product Information Bulletin* and Guidelines [2].
3. Insert tubes into the reaction module of the device.
4. Run the amplification program with fluorescence detection.
5. Analyze results after the amplification program is completed.

## 9. DATA ANALYSIS

Analysis of results is performed by the software of the real-time PCR instrument used by measuring fluorescence signal accumulation in two channels:

- The signal of the *Mycoplasma hominis* DNA amplification product is detected in the channel for the FAM fluorophore;
- The signal of the Internal Control amplification product is detected in the channel for the JOE fluorophore.

Results are interpreted by the crossing (or not-crossing) the fluorescence curve with the threshold line set at the specific level that corresponds to the presence (or absence) of a Ct value of the DNA sample in the corresponding column of the result grid.

Principle of interpretation is the following:

- *Mycoplasma hominis* DNA is **detected** in a sample if the Ct value is determined in the result grid in the channel for the FAM fluorophore. Moreover, the fluorescence curve of the sample should cross the threshold line in the area of typical exponential growth of fluorescence.
- *Mycoplasma hominis* DNA is **not detected** in a sample if the Ct value is not determined (absent) in the result grid in the channel for the FAM fluorophore (the fluorescence curve does not cross the threshold line), whereas the Ct value in the channel for JOE fluorophore is less than the specified boundary Ct value.
- The result is **invalid** if the Ct value is not determined (absent) in the channel for FAM fluorophore, whereas the Ct value in the channel for JOE fluorophore is not determined (absent) or greater than the specified boundary Ct value. In such cases, the PCR analysis should be repeated for such samples.

**NOTE:** Boundary Ct values are specified in the *Important Product Information Bulletin* enclosed to the PCR kit. See also Guidelines [2].

**The result of the analysis is considered reliable only if the results obtained for the Positive and Negative Controls of amplification as well as for the Negative Control of extraction are correct (see Table 3).**

Table 3

Control	Stage for control	Results for controls	
		Ct value in channel for fluorophore	
		FAM	JOE
C–	DNA extraction	Absent	<boundary value
NCA	PCR	Absent	Absent
C+	PCR	<boundary value	<boundary value

## 10. TROUBLESHOOTING

The results of analysis are not taken into account in the following cases:

1. If the Ct value determined for the Positive Control of Amplification (C+) in the channel for the FAM fluorophore is greater than the boundary Ct value or absent, the amplification should be repeated for all samples in which *Mycoplasma hominis* DNA was not detected.
2. If the Ct value is determined for the Negative Control of Amplification (NCA) and/or Negative Control of Extraction (C–) in the channel for the FAM fluorophore, the PCR analysis (beginning with the DNA extraction stage) should be repeated for all samples in which *Mycoplasma hominis* DNA was detected.

If you have any further questions or if encounter problems, please contact our Authorized representative in the European Community.

## 11. TRANSPORTATION

**AmpliSens® Mycoplasma hominis-FRT** PCR kit should be transported at 2–8 °C for no longer than 5 days.

## 12. STABILITY AND STORAGE

All components of the **AmpliSens® Mycoplasma hominis-FRT** PCR kit are to be stored at 2–8 °C when not in use (except for polymerase (TaqF) and PCR-mix-2-FRT).

All components of the **AmpliSens® Mycoplasma hominis-FRT** PCR kit are stable until the expiry date stated on the label. PCR kit **variant FRT-100 F** can be stored without unpacking at 2 to 8 °C for 3 months from the date of manufacture before opening. Once opened, PCR kit **variant FRT-100 F** should be unpacked in accordance with the storage temperatures for each component. The shelf life of reagents before and after the first use is the same, unless otherwise stated.

Polymerase (TaqF) and PCR-mix-2-FRT are to be stored at the temperature from minus 24 to minus 16 °C.

**NOTE:**

PCR-mix-1-FL *Mycoplasma hominis* is to be kept away from light.

<sup>1</sup> For example, Rotor-Gene 3000/Rotor-Gene 6000 (Corbett Research, Australia), Rotor-Gene Q (Qiagen, Germany).

<sup>2</sup> For example, iCycler, iQ5 (Bio-Rad, USA), Mx3000P, Mx3000 (Stratagene, USA).

## 13. SPECIFICATIONS

### 13.1. Sensitivity

The analytical sensitivity of AmpliSens® *Mycoplasma hominis*-FRT PCR kit is specified in the table below.

Clinical material	Transport medium	Nucleic acid extraction kit	PCR kit	Sensitivity, GE/ml <sup>3</sup>
Urogenital swabs	Transport Medium for Swabs or Transport Medium with Mucolytic Agent	DNA-sorb-AM	variant FRT	1x10 <sup>3</sup>
Urine (pretreatment is required)	–	DNA-sorb-AM	variant FRT	2x10 <sup>3</sup>

### 13.2. Specificity

The analytical specificity of AmpliSens® *Mycoplasma hominis*-FRT PCR kit is ensured by selection of specific primers and probes as well as by selection of stringent reaction conditions. The primers and probes were checked for possible homologies to all sequences deposited in gene banks by sequence comparison analysis.

Nonspecific reactions were absent while testing human DNA samples and DNA panel of the following microorganisms: *Ureaplasma urealyticum* and *U. parvum*; *Mycoplasma genitalium*; *Gardnerella vaginalis*; *Lactobacillus* spp.; *Escherichia coli*; *Staphylococcus aureus*; *Streptococcus pyogenes*; *Streptococcus agalactiae*; *Candida albicans*; *Neisseria flava*, *N. subflava*, *N. sicca*, *N. mucosa*, and *N. gonorrhoeae*; *Chlamydia trachomatis*; *Trichomonas vaginalis*; *Treponema pallidum*; *Toxoplasma gondii*; HSV types 1 and 2; CMV; and HPV.

The clinical specificity of AmpliSens® *Mycoplasma hominis*-FRT PCR kit was confirmed in laboratory clinical trials.

## 14. REFERENCES

- Handbook "Sampling, Transportation, and Storage of Clinical Material for PCR Diagnostics", developed by Federal Budget Institute of Science "Central Research Institute for Epidemiology" of Federal Service for Surveillance on Consumers' Rights Protection and Human Well-Being.
- Guidelines "Real-Time PCR Detection of STIs and Other Reproductive Tract Infections", developed by Federal Budget Institution of Science "Central Research Institute for Epidemiology" of Federal Service for Surveillance on Consumers' Rights Protection and Human Well-Being, Moscow.

## 15. QUALITY CONTROL

In accordance with Federal Budget Institute of Science "Central Research Institute for Epidemiology" ISO 13485-Certified Quality Management System, each lot of AmpliSens® *Mycoplasma hominis*-FRT PCR kit has been tested against predetermined specifications to ensure consistent product quality.

### List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
28.06.11 RT	Cover page, text	The name of Institute was changed to Federal Budget Institute of Science "Central Research Institute for Epidemiology"
20.11.15 PM	Through the text	Corrections according to the template
	8.1. DNA extraction	Information about controls of extraction was added
	9. Data analysis 10. Troubleshooting	The sections was rewritten
29.01.18 PM	3. Content	The colour of the reagent was specified
14.03.18 PM	Footer, 3. Content	<b>REF</b> R-B3(iQ)-CE was deleted
14.03.19 TA	2. Principle of PCR detection	The information about the enzyme UDG was added. The information about "hot-start" was corrected
15.05.20 VA	Through the text	The text formatting was changed
	Footer	The phrase "Not for use in the Russian Federation" was added
	2. Principle of PCR detection	The table with targets was added
17.03.21 MA	–	The name, address and contact information for Authorized representative in the European Community was changed
30.11.21 EM	12. Stability and storage	The information about storage conditions for 3 months from the date of manufacture and subsequent unpacking was added
	Through the text	The reference numbers of nucleic acid extraction kits and transport mediums were deleted

## AmpliSens®



Ecoli Dx, s.r.o., Purkyňova 74/2  
110 00 Praha 1, Czech Republic  
Tel.: +420 325 209 912  
Cell: +420 739 802 523



Federal Budget Institute of Science "Central Research Institute for Epidemiology"  
3A Novogireevskaya Street  
Moscow 111123 Russia

<sup>3</sup> Genome equivalents (GE) of the pathogen agent per 1 ml of a sample placed in the transport medium.