

Certificate of Analysis

Page 1

ePure Blood DNA Extraction Kit 200

Date of issue 14. 6. 2021

Certificate of Analysis

Product Name: ePure Blood DNA Extraction Kit 200

Catalogue Number: E2001 **LOT Number: 204623**

Quality Control	Passed
-----------------	--------

Magnetic Beads

The Magnetic bead binding capacity was tested by determined the recovery obtained with 100 bp. and 600 bp. DNA fragment. The recovery should be greater than 80%



Proteinase K activity

One unit is defined as the amount of enzyme that will be liberate 1 μ mol tyrosine from casein per min at 37 °C and pH 7,5



Buffers

Conductivity and pH of buffers were tested to be within the range below

Buffers	Conductivity (mS/cm)	рН
Lysis buffer 1	17 – 25,6 (1/10 dilution)	4,5 – 5,5
Binding buffer 1	1,0 – 2,5	7,5 – 7,9
Wash buffer 1	80 – 120	7,0 – 7,5
Wash buffer 2	35 – 41	7,5 - 8,2
Wash buffer 3	35 – 41	7,5 – 8,2
Elute buffer	0,5 – 0,7	7,8 - 8,3

Carrier RNA

Molecular weight: 100 - 500 kD

2,3 µmol/mg Poly (A) (from absorbance) in relation to one mononucleotide unit. Chromatographically homogeneous.



Passed

Quality Control

Performance

Using fresh whole blood sample to check the difference between lots. Analysis by spectrophotometer and agarose gel electrophoresis, the variation should be below 10% **Storage conditions**



ePure Blood DNA Extraction Kit should be stored at room temperature $(15 - 25 \, ^{\circ}\text{C})$. Storage at higher temperature should be avoided.

Quality Assurance