

## **Certificate of Analysis**

Page 1

# **ePure Blood DNA Extraction Kit 200**

Date of issue 25. 5. 2022

# **Certificate of Analysis**

Product Name: ePure Blood DNA Extraction Kit 200

**Catalogue Number:** E2001 **LOT Number:** 21401

Quality Control	Passed
-----------------	--------

#### **Magnetic Beads**

The Magnetic bead binding capacity was tested by determined the recovery obtained with 100 bp. and 600 bp. DNA fragment. The recovery should be greater than 80%



#### **Proteinase K activity**

One unit is defined as the amount of enzyme that will be liberate 1  $\mu$ mol tyrosine from casein per min at 37 °C and pH 7,5



#### **Buffers**

Conductivity and pH of buffers were tested to be within the range below

Buffers	Conductivity (mS/cm)	рН
Lysis buffer 1	17 – 25,6 (1/10 dilution)	4,5 - 5,5
Binding buffer 1	1,0 – 2,5	7,5 – 7,9
Wash buffer 1	80 – 120	7,0 – 7,5
Wash buffer 2	35 – 41	7,5 - 8,2
Wash buffer 3	35 – 41	7,5 – 8,2
Elute buffer	0,5 – 0,7	7,8 - 8,3

## **Carrier RNA**

Molecular weight: 100 - 500 kD

 $2,\!3~\mu mol/mg$  Poly (A) (from absorbance) in relation to one mononucleotide unit. Chromatographically homogeneous.



**Passed** 

# **Quality Control**

**Performance** 

Using fresh whole blood sample to check the difference between lots. Analysis by spectrophotometer and agarose gel electrophoresis, the variation should be below 10% **Storage conditions** 



ePure Blood DNA Extraction Kit should be stored at room temperature (15 – 25 °C). Storage at higher temperature should be avoided.

Quality Assurance