

Notification of change

Product Name:

Cat.no.E2001 ePure Blood DNA Extraction kit 200 (400ul), (48)

Cat.no.E2003 ePure Viral NA Extraction kit (400ul), (48)

Content of the change:

The expiry date is extended from to "24 months" for catalog numbers: E2001, E2003, lots:204623, 204917C, 204919A.

Dear valued customers,

ePure purification kits have completed the change for stability to 2022.01, a stability study has been performed and the conclusion is that the expiration date can be extended to 24 months.

The kits (lots 204623, 204917C, 204919A) were produced after the evaluation lot for stability study, so these lots have the same process, material and quality system, therefore the expiry date lots 204623, 204917C, 204919A can be extended to 24 months from the date of production.

As confirmed by the internal stability study and under the conditions of proper storage at the temperature specified in the relevant IFU, the stability of the batch below is as follows:


Cat. No.	Product Name	lot	Original expiry date	Extended expiry date
E2001	ePure Blood DNA Extraction kit 200 (400ul), (48)	204623	31.05.22	14.06.23
E2003	ePure Viral NA Extraction kit (400ul), (48)	204917C	30.06.22	15.12.23
E2003	ePure Viral NA Extraction kit (400ul), (48)	204919A	31.12.22	23.02.24

Please forward this information to all relevant persons and other relevant organisations who may need to know about this change.

We apologize for any inconvenience and thank you for your assistance.

For more information, please contact the email address below:

ecoli@ecolidx.com


23.6.22
Ing. Jana Vlastníková
Ecoli Dx, s.r.o.
Purkyňova 74/2,
110 00 Praha 1, Nové Město
IČ: 107 07 409, BIC: CZ10 0707 409
Tel: +420 325 209 912
Head of the division Ecoli Dx

Ecoli Dx s.r.o.
Purkyňova 74/2
110 00 Praha 1

