

Certificate of Analysis

Page 1

ePure STD DNA Extraction Kit

Date of issue 11. 10. 2023

Certificate of Analysis

Product Name: ePure STD DNA Extraction Kit

Catalogue Number: E2007 LOT Number: 231007B

Quality Control Passed

Magnetic Beads

The Magnetic bead binding capacity was tested by whole blood or DNA virus as samples to check the difference between lots, the variation should meet our standards

Proteinase K activity

One unit is defined as the amount of enzyme that will be liberate 1 μ mol tyrosine from casein per min at 37 °C and pH 7,5

Buffers

Conductivity and pH of buffers were tested to be within the range below

Buffers	Conductivity (mS/cm)	рH	Result
Lysis buffer	20 – 35 (1/10 dilution)	4,5-5,5	Passed
Binding buffer	1,0 – 2,5	7,5-7,9	Passed
Wash buffer 1	80 – 120	7,0-7,5	Passed
Wash buffer A	35 – 41	7,5-8,2	Passed
Wash buffer B	35 – 41	7,5-8,2	Passed
Elute buffer	0,5-0,7	7,8-8,3	Passed

Performance

Using liver tissue sample to check the difference between lots. Analysis by spectrophotometer and agarose gel electrophoresis, the variation should below 10%

Storage conditions

ePure STD DNA Extraction Kit should be store at room temperature (15 - 25 °C). Storage at higher temperature should be avoided.

Quality Assurance